



GRIMM

Aerosol Technik

Environmental Dust Monitor EDM180

Eng. Wolfgang Brunnhuber
Spectrometer Division Manager

www.GRIMM-aerosol.com

www.METS-Oman.com



AGENDA

technology

User benefits

Approvals & certifications

This presentation was brought to you by
Mazoon Environmental and Technological Services LLC
www.METS-Oman.com, info@METS-Oman.com

www.GRIMM-aerosol.com



EDM180 stationary ambient monitor


**EN12341 & EN14907
US-EPA
approved**



- * Equivalent to EN12341, EN14907 and US-EPA
- * 19" rack mount
- * Special NAFION dehumidification system
- * GESYTEC MODE for networks
- * Long time measurement
- * Very low TCO's

www.GRIMM-aerosol.com

www.METS-Oman.com



Datum	Name	Version	Alle Rechte vorbehalten © 2007 Alle Rechte vorbehalten © 2007
02.03.08	Meßgerät 180	1.0	
02.03.08	Aufbau Meßgerät 180	1.0	

www.GRIMM-aerosol.com



General features

- 31 channel size distribution
- Different measurements simultaneously
- No probe heating only dehumidification
- No loss of Semi-Volatile compounds
- Nano particles possible



www.GRIMM-aerosol.com

www.METS-Oman.com



GRIMM
Aerosol Technik

NAFION DRYER

www.GRIMM-aerosol.com



Sampling pipe with Nafion tube



www.GRIMM-aerosol.com

www.METS-Oman.com



Benefits of Nafion


- **Nafion transfers moisture quickly**
Water transfers through Nafion as a first order kinetic reaction.
Water transport is driven by partial vapor pressure differential.
- **Nafion transfers water selectively**
Water associates readily with sulfonic acid groups and permeates through the membrane wall.
- **Nafion needs no maintenance**

By the working principle of the Nafion it has the big advantage that it doesn't need any cleaning and has a long life time (if used under normal ambient air conditions over 10years) without loss of performance.
- **No loss of semi volatile compounds**

www.GRIMM-aerosol.com



Dryer flow schematic




www.GRIMM-aerosol.com

www.METS-Oman.com




RAW Data – 15min average values



www.GRIMM-aerosol.com



RAW Data – comparison to reference daily values



www.GRIMM-aerosol.com

www.METS-Oman.com



General features

Results are :

- in real time – fast mode 6 seconds
- not location specific
- no seasonal mass corrections
- no radioactive source
- no sensitivity to vibrations.
- with RS-232 and enviro-protocoll
- data is reproducable
- instruments can be exactly calibrated

www.GRIMM-aerosol.com



Quality control & calibration

www.GRIMM-aerosol.com

www.METS-Oman.com




1. manual & electrical



www.GRIMM-aerosol.com



2. Single channel Calibration



certified monodisperse Latex
model 78-03 to 78-35
Poly-Styrene Latex


www.GRIMM-aerosol.com

www.METS-Oman.com



3. Calibration procedure

The calibration tower, model 7.851 permits powder injection (on demand) of aerosol particles in a wide size range
0.2 μ m - 30 μ m, dolomite dust model 7.856




The operation is fully computer controlled and permits the access of one to three spectrometers in comparison to one reference "mother unit"

www.GRIMM-aerosol.com



4. Single channel adjustment



www.GRIMM-aerosol.com

www.METS-Oman.com



5. Outdoor control procedure



www.GRIMM-aerosol.com



Quality control by user

every 6 month



- leak test
- zero test
- flow control

www.GRIMM-aerosol.com

www.METS-Oman.com



Quality control by user

K-Line

- permanent control of laser quality
- 24hour automatic self test system
- at every system start

www.GRIMM-aerosol.com



Quality control by user

every 12 – 18 months




www.GRIMM-aerosol.com

www.METS-Oman.com



Data is real time available in the WWW



www.GRIMM-aerosol.com

www.GRIMM-aerosol.com

www.METS-Oman.com

AGENDA

technology

User benefits

Approvals & certifications

www.GRIMM-aerosol.com



User benefits

Maintenance

The system is designed in such a way that only a minimum of maintenance is needed.

Every 6 months a system check with a easy to use test kit.

Calibration control every 12 –18

→ Reduction of life time costs

Automatically Check System

A complete automatically self test and control system is watching over the instruments quality and reports errors

www.GRIMM-aerosol.com

www.METS-Oman.com



User Benefits

Consumables

No Consumables are needed (for example filter band)

→Reduction of costs

(at other instruments this could cost up to 3.000€ within 3 years)

Energy costs are lower

The maximum power consum is 50W/h

→ Reduction of energy costs

(example of Helsinki: 3 EDM180 – 150W, for the same number of values PM10 & PM2.5 – 6 TEOM – 3000W.

Cost save in Helsinki within 10 years 26.000€)

www.GRIMM-aerosol.com



User Benefits

Simultaneous three PM values

The only available fine dust monitor, EN approved for PM-10 with the ability to measure simultaneously PM-10 and PM-2.5 and PM-1 in $\mu\text{g}/\text{m}^3$

Count Values

Additional information of the total fine dust particle concentration will allow the understanding of the aerosol source measured. Data are in Particles / litre.

Risk Management

Fast data presentation, combined with the meteorological sensors will allow to identify the source direction and the possible risk associated. Distribution models can be feed with such information's.

www.GRIMM-aerosol.com

www.METS-Oman.com



Differences to the competitors

(TEOM, SHARP, BAM,...)

- ☞ 4 different values simultaneously in real time
- ☞ No heating of the probe – therefore no loss of SEMI Volatiles – the only instrument that includes the SVC fraction in all PM values
- ☞ No radioactive source is in use
- ☞ No consumables (no change of filters, cartridges,...)
- ☞ No problems with sand storms or desert areas
- ☞ No external pump systems No specification of the measurement range by the probe pipe head (TSP head is used)
- ☞ Easily change of measurement if regulation changes by firmware update – no need of hardware change!

www.GRIMM-aerosol.com



STEP1 - EN12341 – PM10



IV.
Eignung von Messeinrichtungen
zur kontinuierlichen Überwachung von Immissionen

Unter Bezugnahme auf die Nummer 3.2 der Bekanntmachung der für die Durchführung der Richtlinie 96/62/EG des Rates vom 27. September 1996 über die Beurteilung und Kontrolle der Luftqualität zuständigen Stellen vom 1. Oktober 1996 (BAnz. S. 15 126) wird im Auftrag des BMU die Eignung der folgenden Messeinrichtung bekannt gegeben:

1 Schwebstaub (PM₁₀-Fraktion)

1.1 Environmental Dust Monitor Modell 180

Hersteller:
Grimm Aerosol Technik GmbH & Co. KG, 83404 Ainring


- Official publicated certification on the 8th of April 2006 by the German UMEG
- First time that a light scattering instrument passed EN standards for environmental monitoring

www.GRIMM-aerosol.com

www.METS-Oman.com



PM10 – Germany 106 days



Winter time comparison between Grimm and grafimatrical reference system Leckl filter collector
Correlation 0,974 and data quality 95%

www.GRIMM-aerosol.com



EN14907 – PM2.5

A LONG STORY
WITH A FINAL
APPROVED
END


www.GRIMM-aerosol.com

www.METS-Oman.com



PM2.5 – Netherland

Raw data without any correction – 30days,
correlation to reference $R^2 = 0,95$, data quality 100%




www.GRIMM-aerosol.com



PM2.5 – Austria

Raw data without any correction – 34days,
correlation to reference $R^2 = 0,93$, data quality 98%




www.GRIMM-aerosol.com

www.METS-Oman.com



PM2.5 – Germany – 116 days

Corrected data – 116days,
correlation to reference $R^2 = 0,95$, data quality 100%



www.GRIMM-aerosol.com



PM2.5 – Germany – 116 days


UNCORRECTED DATA		INTERCEPT CORRECTION	
REGRESSION OUTPUT			
REGRESSION OUTPUT		REGRESSION OUTPUT	
slope b	1,28 significant	slope b	1,28 significant
uncertainty of b	0,02	uncertainty of b	0,02
intercept a	2,14 significant	intercept a	0,00 not significant
uncertainty of a	0,36	uncertainty of a	0,36
EQUIVALENCE TEST RESULTS			
random term	1,53	random term	1,57
bias at LV	9,07	bias at LV	6,94
combined uncertainty	9,16	combined uncertainty	7,06
exp. rel. comb. uncertainty (%) at LV	73,26 fail	exp. rel. comb. uncertainty (%) at LV	56,40 fail
RM between-sampler uncertainty	1,50	RM between-sampler uncertainty	1,50
SLOPE CORRECTION		INTERCEPT AND SLOPE CORRECTION	
REGRESSION OUTPUT			
REGRESSION OUTPUT		REGRESSION OUTPUT	
slope b	1,00 not significant	slope b	1,00 not significant
uncertainty of b	0,02	uncertainty of b	0,02
intercept a	1,72 significant	intercept a	0,06 not significant
uncertainty of a	0,28	uncertainty of a	0,28
EQUIVALENCE TEST RESULTS			
random term	0,90	random term	0,97
bias at LV	1,63	bias at LV	0,94
combined uncertainty	1,89	combined uncertainty	0,98
exp. rel. comb. uncertainty (%) at LV	15,12 pass	exp. rel. comb. uncertainty (%) at LV	7,87 pass
RM between-sampler uncertainty	1,50	RM between-sampler uncertainty	1,50

www.GRIMM-aerosol.com

www.METS-Oman.com




Comparison PM 2.5 Grimm – TEOM FDMS in Netherland by the DCMR



www.GRIMM-aerosol.com



PM2.5 – Turkey



2 EDM180 compared over 90days

www.GRIMM-aerosol.com

www.METS-Oman.com




EN14907 – PM2.5

A LONG STORY
WITH A FINAL
APPROVED
END

www.GRIMM-aerosol.com



Step 2 - EN14907 – PM2.5



umweltbundesamt^U

Umweltbundesamt GmbH
Spittelauer Lände 5
1090 Wien/Österreich

Der kontinuierliche Staubmonitor GRIMM Modell 180 wurde entsprechend der Vorgaben des Leitfadens „Demonstration of Equivalence of Ambient Air Monitoring Methods“ an vier Messstellen in Österreich gegen die Referenzmethoden für PM₁₀ und PM_{2,5} getestet und die dabei gewonnenen Messdaten nach den Regeln des Leitfadens ausgewertet.

Die Äquivalenz des GRIMM Modell 180 konnte sowohl für PM₁₀ als auch für PM_{2,5} nachgewiesen werden.

www.GRIMM-aerosol.com

www.METS-Oman.com



Winter 2009 - EN14907 – PM2.5

PROJEKTION FÜR umweltbundesamt^U
UMWELT & GEMEINSCHAFT

Zusammenfassung der Ergebnisse des Äquivalenztests GRIMM Modell 180

Der kontinuierliche Staubmonitor GRIMM Modell 180 wurde entsprechend der Vorgaben des Leitfadens „Demonstration of Equivalence of Ambient Air Monitoring Methods“ an vier Messstellen in Österreich gegen die Referenzmethoden für PM₁₀ und PM_{2,5} getestet und die dabei gewonnenen Messdaten nach den Regeln des Leitfadens ausgewertet.

Die Äquivalenz des GRIMM Modell 180 konnte sowohl für PM₁₀ als auch für PM_{2,5} nachgewiesen werden.

Die Messungen wurden an folgenden vier Messstellen durchgeführt:

- Messstelle 1: Graz Süd (städtischer Hintergrund, hohes Konzentrationsniveau)
Dezember 2007 – März 2008
- Messstelle 2: Steyregg (loker verbautes Wohngebiet mit Industrieinfluss)
Juni 2008 – August 2008
- Messstelle 3: Wiesendfeld bei Hohenbrunn (ländlich)
Dezember 2008 – März 2009
- Messstelle 4: Klagenfurt (städtischer Hintergrund, niedriges Konzentrationsniveau)
Juni 2009 – August 2009

Sowohl für PM₁₀ als auch für PM_{2,5} ist die Anwendung einer Kalibrierfunktion erforderlich.

Für PM₁₀ beträgt die Kalibrierfunktion

$$C_{\text{Equivalent}} = C_{\text{Candidate}} \cdot (C_{\text{Corr}} + 0,37) / 1,155$$

Die maximale kombinierte Messunsicherheit wurde an der Messstelle Wiesendfeld beobachtet und betrug 9,6%.

Für PM_{2,5} beträgt die Kalibrierfunktion

$$C_{\text{Equivalent}} = C_{\text{Candidate}} \cdot (C_{\text{Corr}} - 3,3) / 1,065$$

Die maximale kombinierte Messunsicherheit wurde an der Messstelle Wiesendfeld beobachtet und betrug 12,2%.

Marina Franz
Mag. Marina Franz
Stv. Abteilungsleiterin Luftqualität & Energie
Tel.: +43(0)1-313 945862
Fax: +43(0)1-313 945860
E-Mail: marina.boehlich@umweltbundesamt.at

www.GRIMM-aerosol.com



US-EPA

Applicant:		GRIMM USA					
Candidate method:		Model 180 - PM2.5 Class III					
Test site:		BAKERSFIELD, CALIFORNIA - (Site location A)					
Data sets		Number					
Valid data sets available:		46					
Number of valid data sets required for Class III, location A:		46					
Number of valid data sets for this test is:		OK					
Additional data sets needed:		---					
Precision	Data set mean, $\mu\text{g}/\text{m}^3$		Data set precision, $\mu\text{g}/\text{m}^3$		Relative precision (CV)		
		FRM	Candidate	FRM	Candidate	FRM	Candidate
	Mean:	24.9	24.2	0.8	0.4	3.5%	1.7%
	Maximum:	75.1	74.6	2.8	6.3	9.4%	11.3%
	Minimum:	5.3	5.1	0.1	0.1	0.6%	0.1%
	Candidate / FRM Ratio:		97.4%		45.4%		49.1%
RMS Relative Precision for this site:					3.8%	2.6%	
Test requirements - PM2.5 Class III:					10.0%	15.0%	
Precision Test Results for site:					OK	PASS	
Regression statistics		Slope ¹	Intercept ²	Correlation (r)			
Statistics for this test site:		0.977	-0.089	0.99059			
Limits for	Upper:	1.100	2.000				
	Lower:	0.900	-1.873	0.95000			
Test Results (Pass/Fail):		PASS	PASS	PASS			

* Finished fall 2009

www.GRIMM-aerosol.com

www.METS-Oman.com



GRIMM Environmental Dust Monitor EDM180



- 2006 for PM10 according to EN12341
- 2009 for PM10 new EN regulations
- 2009 for PM2.5 according to EN12341
- 2010 US-EPA papers are published
- 2010 Russian Approval finished in summer

www.GRIMM-aerosol.com



Where is it in use?

Generally the EDM180 is running in several states of Europe and the world in national and local authority networks and shows very good results in equivalency to the gravimetric reference methods.

Some of these countries are:

Germany	Bosnia
Belgium	Croatia
Austria	Switzerland
Serbia	Korea
Spain	United Arab Emirates
Czech Republic	Turkey
China	Australia
South Africa	Saudi Arabia
Greece	Canada
Hungary	Netherlands
Mongolia	

www.GRIMM-aerosol.com

www.METS-Oman.com